

MEASUREMENT AND TEST EQUIPMENT

03/21/18

PAGE 1-2	MEASUREMENT AND TEST EQUIPMENT (M&TE) STANDARDS, NIST TRACEABLE NUMBERS AND M&TE DUE DATES – 2 year cycle
PAGE 3	MEASUREMENT AND TEST EQUIPMENT (M&TE) STANDARDS, NIST TRACEABLE NUMBERS AND M&TE DUE DATES – 4 year cycle
PAGE 4-5	MEASUREMENT AND TEST EQUIPMENT (M&TE) STANDARDS, NIST TRACEABLE NUMBERS AND M&TE DUE DATES – 1 year cycle

Pi Tape Texas, LLC.
10235 Robinson Drive
Tyler, TX 75703
USA

Phone # 903-266-9204
Fax # 903-939-0909

PI TAPE® precision measuring products
Quality since 1944



MEASUREMENT AND TEST EQUIPMENT (M&TE) STANDARDS

CALIBRATION OF M&TE'S IS TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS AND TECHNOLOGY. ALL STANDARDS ARE CALIBRATED AT A MINIMUM OF A 4:1 RATIO. (NOTE: GAGE SIZES ARE USED AS REFERENCE SIZES AND ARE NOT A DIRECT CONVERSION FROM INCH TO METRIC OR FROM O.D TO O.C.)

THE FOLLOWING STANDARDS ARE CALIBRATED EVERY **TWO YEARS** TO GUARANTEE ACCURACY:

<u>MT&E</u>	<u>VENDOR</u>	<u>NIST TRACEABLE #</u>	<u>M&TE DUE DATE</u>
3" (76mm) O.D. GAGE # 3000 (9"OC)	AMER GAGE	#106419-1	Sep-27-18
5" (127mm) I.D. RING GAGE # 5000	AMER GAGE	#107813-1	Feb-20-19
12" (300mm) I.D. RING GAGE # 1944	AMER GAGE	#111333-2	Mar-13-20
16" (400mm) I.D. RING GAGE # 6000	TRESCAL	#3491460002	Apr-26-18
24" (600mm) O.D. GAGE #201 (75"OC)	AMER GAGE	#108288-1	Apr-04-19
27" (700mm) O.D. RING GAGE #1128 (84"OC)	AMER GAGE	#108288-2	Apr-04-19
36" (900mm) O.D. RING GAGE #36585 (115"OC)	AMER GAGE	#108288-3	Apr-04-19
44" (1100mm) O.D. RING GAGE #1144 (138"OC)	AMER GAGE	#108288-4	Apr-04-19
WEBBER 6" GAGE BLOCK # 100-4	STARRETT/WEBBER	#16-31539-C	Jun-18-18
WEBBER 6" GAGE BLOCK # 100-2	STARRETT/WEBBER	#16-31539-A	Jun-18-18
WEBBER 6" GAGE BLOCK # 100-6	STARRETT/WEBBER	#16-31539-E	Jun-18-18
WEBBER 6" GAGE BLOCK # 100-7	STARRETT/WEBBER	#16-31539-F	Jun-18-18
WEBBER 6" GAGE BLOCK # 100-5	STARRETT/WEBBER	#16-31539-D	Jun-18-18
WEBBER 6" GAGE BLOCK # 100-3	STARRETT/WEBBER	#16-31539-B	Jun-18-18

THE FOLLOWING STANDARDS ARE CALIBRATED EVERY TWO YEARS TO GUARANTEE ACCURACY:CONTINUE

T	LINEAR MACHINE # 62494	PI TAPE	#62494081116	Aug-11-18
T	LINEAR MACHINE # 113004	PI TAPE	#113004112017	Nov-20-19
T	LINEAR MACHINE #31738	PI TAPE	#31738081116	Aug-11-18
	LINEAR MACHINE #73165	PI TAPE	#73165112016	Nov-17-18
	FRACTIONAL WEBBER GAGE BLOCKS SERIAL # 4.3064SA1	STARRETT/WEBBER	#16-31539-G	Jun-16-18
	WEBBER GAGE BLOCKS SERIAL # R1R005	STARRETT/WEBBER	#17-36194-A	Apr-26-19

T signifies Tamper Resistant Seal – All seals verified and recorded each workday

THE FOLLOWING STANDARDS ARE CALIBRATED EVERY **FOUR YEARS** TO GUARANTEE ACCURACY:

<u>MT&E</u>	<u>VENDOR</u>	<u>NIST TRACEABLE #</u>	<u>M&TE DUE DATE</u>
1-2" MICROMETER # 216	AMER GAGE	#97687-4	Apr-24-18
48" CALIPER #333	TRESCAL	#3499790001	May-03-20

THE FOLLOWING STANDARDS ARE CALIBRATED EVERY ONE YEAR TO GUARANTEE ACCURACY:

<u>MT&E</u>	<u>VENDOR</u>	<u>NIST TRACEABLE #</u>	<u>M&TE DUE DATE</u>
.375" O.D. GAGE # 42796	AMER GAGE	#111060-1	Feb-09-19
.500" O.D. GAGE # 42596	AMER GAGE	#111060-2	Feb-09-19
.625" O.D. GAGE # 42696	AMER GAGE	#111060-3	Feb-09-19
1" (25mm) O.D. GAGE #1000 (3"OC)	AMER GAGE	#111060-4	Feb-09-19
1" (25mm) O.D. GAGE #101	AMER GAGE	#109567-1	Aug-17-18
4.9" (125mm) O.D. GAGE #102098 (15"OC)	AMER GAGE	#111333-3	Mar-13-19
5" (127mm) O.D. GAGE #555 (16"OC)	AMER GAGE	#110793-2	Jan-08-19
9" (228mm) O.D. GAGE #900 (28"OC)	AMER GAGE	#109086-1	Jul-03-18
13" (315mm) O.D. GAGE #042299 (40"OC)	AMER GAGE	#110793-3	Jan-08-19
19" (480mm) O.D. GAGE #1619 (59"OC)	TRESCAL	#3862700001	Apr-18-18
20" (500mm) O.D. GAGE #042199 (62"OC)	TRESCAL	#3952860001	Jul-11-18
DIGITAL THERMOMETER #52590	AMER GAGE	#111333-4	Mar-13-19
0-10# SPRING SCALE # 132040	AMER GAGE	#109567-2	Aug-17-18

T signifies Tamper Resistant Seal – All seals verified and recorded each workday

THE FOLLOWING STANDARDS ARE CALIBRATED EVERY ONE YEAR TO GUARANTEE ACCURACY:(CONTINUE)

T	0-30# SPRING SCALE # 52490	AMER GAGE	#110793-1	Jan-08-19
	1# WEIGHT #6679-1	AMER GAGE	#109567-3	Aug-17-18
	2# WEIGHT # 2S7E	AMER GAGE	#109567-3	Aug-17-18
	5# WEIGHT #6676-1	AMER GAGE	#109567-3	Aug-17-18
T	HUMIDITY INDICATOR #24826	AMER GAGE	#111333-1	Mar-13-19
	OPTICAL COMPARATOR #R041507	TRESCAL	#4006309501	Oct-04-18
	INFORMER II DIGITAL THERMOMETER #91798	AMER GAGE	#108183-1	Mar-28-18
	0-39" LINEAR TAPE STRIP SN L08171732 / BASE SN L08171711	PI TAPE	#L08171732092517	Sep-25-18

T signifies Tamper Resistant Seal – All seals verified and recorded each workday

End of document